

INSTRUCTIONS FOR USE

ParaSure Self-Adhesive Class 6 Indicators are designed for use in steam sterilizers and are specifically calibrated to the individual cycles listed below. When used as directed, the Self-Adhesive Indicator gives visual confirmation that sterilization conditions were met at the location in the sterilizer chamber.

COLOUR CHANGE

During steam sterilization, the indicator changes from violet to teal.

INSTRUCTIONS FOR USE

1. Place a ParaSure Self-Adhesive Class 6 Indicator into each pack, peel-pouch or tray to be processed.
2. Process the loads according to the sterilizer manufacturer's instructions.
3. Following the sterilization cycle, visually verify that the indicator spot has changed colour from violet to teal.
4. If the indicator spot has not changed colour, it suggests inadequate sterilization conditions and the items must be re-sterilized with new indicator monitors.

SAFETY PRECAUTIONS

- *If there is any doubt about sterility of item, it must be considered NOT sterile*
- *Do not use any strips whose indicator box is not blue in colour prior to exposure.*
- *Not recommended for use in a "flash" sterilization cycle.*

STORAGE

ParaSure Self-Adhesive Class 6 Indicators should be stored between 50°F-100°F (10°C-38°C) in the re-sealable bag provided. No special storage conditions are necessary after exposure to steam sterilization conditions.

EXPIRY DATE

Expiry date is printed on the product packaging.

LOT NUMBER

A unique identification code, LOT printed on each indicator strip and packaging labelling.

CALIBRATIONS

PS1210 3.5 minutes at 134°C

PS1230 5.3 minutes at 134°C

PS1220 4 minutes at 134°C

PS1240 7 minutes at 134°C