

## Load Control Indicator User Guide



PS1400 EN 867-5 Helix Device  
PS1410 3.5 mins. at 134°C Load Control  
PS1420 4.0 mins. at 134°C/12 mins. at 121°C Load Control

PS1430 5.3 mins. at 134°C/15 mins. at 121°C Load Control  
PS1440 7.0 mins. at 134°C/20 mins. at 121°C Load Control  
PS1450 18.0 minute Load Control



Unprocessed indicator




Successfully processed indicator

**PASS**  
Items in tray/pack  
can be safely used



Unsuccessfully processed indicator

**FAIL**  
Do NOT use items  
in tray/pack



Unsuccessfully processed indicator

**FAIL**  
Do NOT use items  
in tray/pack



Unsuccessfully processed indicator

**FAIL**  
Do NOT use items  
in tray/pack

